



Notes on Using “Ni Setup Tool”

- Upon first use of the Ni Setup Tool, always perform microscope system settings with the setup function, and transmit this information to the microscope system using “SEND.”
- When setup information is transmitted to the microscope system, this will overwrite the previous information held in its memory. We recommend that the information, including arbitrary registrations set up with the setup function or other setting function, be assigned a filename and saved on the PC side, after transmission.